U2000 Series USB Power Sensors

The compact, low-cost alternative to conventional power measurement solutions





Table of Contents

Why Keysight's Power Meters and Sensors?	3
Compact Solutions for Testing Today's RF and Microwave Communication Systems	4
Introducing the U2000 Series USB Power Sensors	5
Take a Closer Look	8
Specifications	9
Using the U2000 Series with the BenchVue Software	26
System and Installation Requirements	27
Ordering Information	28



Why Keysight's Power Meters and Sensors?

Reliable, high-performing solutions

Every power meter and sensor from Keysight Technologies, Inc. consistently delivers great results.

A sure investment for many years to come

Code-compatibility between power meters reduces the need for re-coding. Not only that, all Keysight power meters are backward- compatible with most legacy power sensors.

One specific application: One right solution

Keysight offers a wide selection of power meters and sensors for practically all application needs wireless communications, radar pulse measurements, component test, and more.

Global network support

No matter where you are, Keysight is committed to giving you the 24-hour support you need regarding our products, applications, or services.

Keysight's power meters have long been recognized as the industry standard for RF and microwave power measurements.



Compact Solutions for Testing Today's RF and Microwave Communication Systems

For installation and maintenance of base stations

- Lightweight and rugged
- Simple set-up and usage
- Portable with low power consumption
- Wide dynamic and frequency ranges
- Quick and easy testing with large display of readings
- Internal trigger eliminates the problem for applications that do not have external trigger signal
- Trace display capability enables easy gate setup

For production testing of wireless components

- Compact build saves rack space
- Simple set-up and usage
- Wide dynamic and frequency ranges
- Fast reading speed
- Internal zeroing reduces test time and sensor wear-and-tear
- Quick and easy multiple channels testing with simultaneous display of readings, limits and alerts
- Seamless integration to system with industry standard SCPI
- Internal trigger eliminates the problem for tests that do not have external trigger signal
- Trace display capability enables easy trigger level and gate setup

For R&D of wireless components

- Compact build saves bench space
- Simple set-up and usage
- Wide dynamic and frequency ranges
- High accuracy
- Advanced troubleshooting of designs with simultaneous display of multiple readings, measurement math and data recording
- Internal trigger eliminates the problem for applications that do not have external trigger signal



Introducing the U2000 Series USB Power Sensors

The U2000 Series enable simpler, lower-cost power measurements versus conventional power meter and sensor combinations. Now with nine high-performance models, the U2000 Series USB power sensors offer compact, high-performance solutions for today's CW and modulated signals.



Features

- Compact, lightweight solutions
- Quick, simple set-up
- High accuracy, high power
- Internal zeroing capability
- Fast reading speed
- Wide frequency range:
- 9 kHz to 24 GHz (sensor option dependent)
- Wide dynamic range:
- -60 to +44 dBm
- Support internal and external trigger measurements ¹
- Trace display capability enables easy trigger level and gate setup for burst signals
- Allows remote measurements beyond cable length
- Enables monitoring of more than 20 channels simultaneously
- Converts select Keysight instruments to power meters
- Feature-packed software provides various capabilities for easy testing and analysis
- Average power measurements of CW and modulated signals, including GSM, EDGE, WLAN and WiMAX[™]

¹ Except U2004A model



Compact "power meters", simple set-up

The U2000 Series are standalone sensors. That means they essentially operate like power meters, just in smaller forms. No reference calibrator is required. The fact that each sensor draws minimal power from a USB port — and that it doesn't need additional triggering modules or power adaptors to operate — makes it more portable, especially for base station testing. Setting up is easy: just plug it to the USB port of your PC or laptop — or even select network or handheld spectrum analyzer — and start your power measurements. The figure below illustrates the very simple, straightforward setup of the U2000 Series.



High accuracy

Each U2000 Series sensor provides excellent linearity, SWR and uncertainty specifications, so you can be confident in every measurement you make.

Wide range, high power

The U2000 Series' dynamic range spans across a wide 80 dB, taking on high power up to +44 dBm.

Remote monitoring and tests

With the U2000 Series sensor plugged to a networked USB hub, you can conveniently monitor power measurements of an antenna tower from the control room, beyond the limits of USB cable lengths.

Faster production testing of multiple channels

The U2000 Series' fast measurement speed helps reduce test time. This, coupled with the capability to enable monitoring of more than 20 channels simultaneously, is an advantage in the production line where efficiency is of utmost priority.



The U2000 Series has both internal and external zeroing capabilities. With internal zeroing, high isolation switches in the sensor are opened to isolate the sensor from the device-under-test (DUT) it is connected to. As such, you don't need to power-off the DUT or disconnect the sensors. This speeds up testing and reduces sensor wear-and-tear.

No manual input of calibration data is required. All calibration factors, as well as temperature and linearity corrections, are stored in the sensors' EEPROM, auto- downloaded at calibration.

Often, you'd need to automate your tests. The U2000 Series sensors are well-equipped for seamless integration to your system with industry standard SCPI compatibility. They also come with built- in triggering capability to allow receipt of external triggers from other instruments.

Transform your signal generators and spectrum analyzers into accurate power meters

You could literally have a power meter next to you — or instead, turn your Keysight signal generators, handheld spectrum analyzers or FieldFox handheld RF/Microware analyzers into a power meter for accurate power measurements. Even with the U2000 connected, you can switch between power measurements and the device's original function at any time. You can also use the U2000 with your Keysight network analyzers for source power calibration.





Intuitive power analysis software

The U2000 Series is supported by the Keysight BenchVue software and BV0007B Power Meter/Sensor Control and Analysis app. Once you plug the USB power sensor into a PC and run the software you can see measurement results in a wide array of display formats and log data without any programming.

For more information, www.keysight.com/find/BenchVue



Take a Closer Look



Diode-based sensors frequently rely on the application of correction factors to extend their dynamic range beyond their square-law region, typically in the range of -70 to -20 dBm. While this technique achieves measurement of CW signals over a wide dynamic range, it fails to do so for modulated signals when the signal level is above the square-law region. Modulated signals must be padded down, with their average and peak power levels within the diode square-law region, for accurate average power measurement.

The U2000 Series USB power sensors are true-average, wide-dynamic-range RF/ microwave power sensors, based on a dual-sensor diode pair/attenuator/diode pair topology as proposed by Szente et. al. in 1990.²

² US Patent #4943764 assigned to Hewlett-Packard Company



The simplified block diagram shown here illustrates this technique.



This technique ensures diodes in the selected signal path are kept in their square law region — with output current and voltage proportional to input power. The diode pair/attenuator/diode pair assembly can then yield the average of complex modulation formats across a wide dynamic range, irrespective of signal bandwidth.

The dual range Modified Barrier Integrated Diode (MBID)³ package incorporates diode stacks in place of single diodes. This further improves measurement accuracy of high-level signals with high crest factors without incurring damage⁴ to the sensor.

Implementation of both techniques in the U2000 Series USB sensors enable effective average power measurements of a wide range of signals, including multitone and spread spectrum signals used in CDMA, W-CDMA and digital television systems.

Specifications

Specifications contained in this chapter are valid ONLY after proper calibration of the power sensor and apply to continuous wave (CW) signals unless otherwise stated. The recommended calibration interval for this product is one year.

Specifications apply over a temperature range 0 to +55 °C unless otherwise stated. Specifications quoted over a temperature range of 25 °C \pm 10 °C apply to a relative humidity of 15% to 75% and conform to the standard environmental test conditions. Specifications are valid after a 30-minute warm-up period.

Supplemental characteristics, shown in italics, are intended to provide useful information with regard to applying the power sensors in that they contain typical, but non-warranted performance parameters. These characteristics are shown in italics or denoted as "typical", "nominal" or "approximate".

⁴ Refer to "Maximum Power" on page 9 for maximum power handling specifications.



³ November 1986 Hewlett-Packard Journal pages 14-2, "Diode Integrated Circuits for Millimeter-Wave Applications".

Measurement speed

Normal: 20 readings/s

x2: 40 readings/s

Fast: 110 readings/s

Buffered (50 readings): 1000 readings/s ⁵

The U2000 Series USB sensors have two measurement modes:

Average only mode (default mode): Optimized for wide dynamic range. In this measurement mode, a trigger can be controlled externally via TTL input.

Normal⁶ mode: Used for making average power measurement in a defined time interval (time-gated measurement) with reduced dynamic range. A trigger can be derived from an RF signal (internal trigger) or controlled externally via TTL input (external trigger).

Model Frequency range		Power range	Maximum power		
U2000A	10 MHz to 18 GHz	–60 to +20 dBm	+25 dBm avg, 20 VDC		
U2001A	10 MHz to 6 GHz		+33 dBm pk, < 10 μs		
U2002A	50 MHz to 24 GHz				
U2004A	9 kHz to 6 GHz	-60 to +20 dBm	+25 dBm avg, 5 VDC		
			+33 dBm pk, < 10 μs		
U2000B	10 MHz to 18 GHz	-30 to +44 dBm	+45 dBm avg, 20 VDC		
U2001B	10 MHz to 6 GHz		+47 dBm pk, 1 μs		
U2000H	10 MHz to 18 GHz	-50 to +30 dBm	+33 dBm avg, 20 VDC		
U2001H	10 MHz to 6 GHz		+50 dBm pk, 1 μs		
U2002H	50 MHz to 24 GHz	-50 to +30 dBm	+33 dBm avg, 10 VDC		
			+50 dBm pk 1 us		

Frequency and power ranges

+50 dBm pk, 1 μs

⁶ Not applicable for U2004A.



⁵ The 1000 reading/s is the derived measurement speed from the first 50 readings in buffered mode. The maximum number of measurements that can be obtained in one second is 250 readings in buffered mode.

Maximum SWR

Model	Frequency range	Maximum SWR (25 °C ± 10 °C)	Maximum SWR (0 °C to 55 °C)
U2000A	10 to 30 MHz	1.15	1.21
	> 30 MHz to < 2 GHz	1.13	1.15
	2 to < 14 GHz	1.19	1.20
	14 to < 16 GHz	1.22	1.23
	16 to 18 GHz	1.26	1.27
U2001A	10 to 30 MHz	1.15	1.21
	> 30 MHz to < 2 GHz	1.13	1.15
	2 to 6 GHz	1.19	1.20
U2002A	50 MHz to 2 GHz	1.13	1.15
	> 2 to 14 GHz	1.19	1.20
	> 14 to 16 GHz	1.22	1.23
	> 16 to 18 GHz	1.26	1.27
	> 18 to 24 GHz	1.30	1.30
U2004A	9 kHz to 2 GHz	1.13	1.15
	> 2 to 6 GHz	1.19	1.20
U2000B	10 MHz to 2 GHz	1.12	1.14
	> 2 to < 12.4 GHz	1.17	1.18
	12.4 to 18 GHz	1.24	1.25
U2001B	10 MHz to 2 GHz	1.12	1.14
	> 2 to 6 GHz	1.17	1.18
U2000H	10 MHz to < 8 GHz	1.15	1.17
	8 to < 12.4 GHz	1.25	1.26
	12.4 to 18 GHz	1.28	1.29
U2001H	10 MHz to 6 GHz	1.15	1.17
U2002H	50 MHz to < 8 GHz	1.15	1.17
	8 to < 12.4 GHz	1.25	1.26
	12.4 to 18 GHz	1.28	1.29
	> 18 to 24 GHz	1.30	1.31



Typical SWR for U2000A sensor (25 °C ± 10 °C)



Typical SWR for U2001A sensor (25 °C ± 10 °C)



Typical SWR for U2002A sensor (25 °C ± 10 °C)





Typical SWR for U2004A sensor (25 °C ± 10 °C)



Typical SWR for U2002H sensor (25 °C ± 10 °C)



Typical SWR for U2000H sensor (25 °C ± 10 °C)





Typical SWR for U2000B sensor (25 °C ± 10 °C)



Typical SWR for U2001H sensor (25 °C ± 10 °C)



Typical SWR for U2001B sensor (25 °C ± 10 °C)





Switching point

The U2000 Series power sensors have two measurement paths: a low-power path and a high-power path, as shown in the table below.

Models	AUTO (default) range	Low power path	High power path	Switching point
U2000/1/2/4A	–60 to +20 dBm	–60 to –7 dBm	-7 to +20 dBm	–7 dBm
U2000/1/2H	–50 to +30 dBm	–50 to +3 dBm	+3 to +30 dBm	+3 dBm
U2000/1B	-30 to +44 dBm	-30 to +23 dBm	+23 to +44 dBm	+23 dBm

Each power sensor automatically selects the proper power level path. To avoid unnecessary switching when the power level is close to the switching point, switching point hysteresis has been added.

Offset at switching point: $\leqslant \pm$ 0.5% ($\leqslant \pm$ 0.02 dB) typical

Switching point hysteresis: ± 0.5 dBm typical

Example with U2000 "A" suffix sensors: Switching point for the U2000/1/2/4A sensors is at -7 dBm. Hysteresis causes the low power path to remain selected until approximately -6.5 dBm as the power level is increased. Above this power, the high-power path is selected. The high-power path remains selected until approximately -7.5 dBm is reached as the signal level decreases. Below this power, the low power path is selected.

Power accuracy

Average only mode power accuracy⁷ (with exclusions).

Model	Power range	Accuracy ¹ (25 °C ± 10 °C)	Accuracy ¹ (0 °C to 55 °C)
U2000/1/2/4A	–60 to +20 dBm	± 3.0%	± 3.5%
U2000/1/2H	–50 to +30 dBm	± 4.0%	± 5.0%
U2000/1B	-30 to +44 dBm	± 3.5%	± 4.0%

Specifications valid with the following conditions:

- After zeroing⁸
- Number of averages = 1024
- After 30 minutes of power-on warm-up

⁸ It is advisable to perform external zeroing on the U2000 Series power sensor for power measurement level below -30 dBm. During the external zeroing process, the RF input signal must be switched off or the device-under-test disconnected from the U2000 Series power sensor.



⁷ This accuracy is essentially a combination of linearity, instrumentation accuracy, and traceability to absolute accuracy at 50 MHz, 0 dBm. Note: Mismatch uncertainty, calibration factor uncertainty, and power level dependent terms (zero set, drift, and noise) are excluded in this specification and specified elsewhere in the data sheet.

Normal mode power accuracy^{9, 10}, (with exclusions).

Model	Power range	Accuracy (25 °C ± 10 °C)
U2000/1/2/A	–30 to +20 dBm	± 4.0%
U2000/1/2H	-20 to +30 dBm	± 5.0%
U2000/1B	0 to +44 dBm	± 4.5%

Typical power accuracy at 25 °C for U2000/1/2/4A sensors^{11, 12}



Typical power accuracy at 25 °C for U2000/1B sensors^{11, 12}



⁹ This accuracy is essentially a combination of linearity, instrumentation accuracy, and traceability to absolute accuracy at 50 MHz, 0 dBm. Note: Mismatch uncertainty, calibration factor uncertainty, and power level dependent terms (zero set, drift, and noise) are excluded in this specification and specified elsewhere in the data sheet.

¹² After zeroing, 30 minutes of power-on warm-up, and 1024 averages.



¹⁰ The accuracy for -7 to +1 dBm (U2000/1/2A), +3 to +11 dBm (U2000/1/2H), and +23 to +31 dBm (U2000/1B) power level will be dominated by zero set and measurement noise. For overall accuracy, refer to the measurement uncertainty calculator which is available on the Keysight Technologies Web site.

¹¹ Measurement uncertainty ≤ 1.9%. At room temperature and excluding power level dependent terms (zero set, drift, and noise). Refer to Keysight Fundamentals of RF and Microwave Power Measurements (Part 3) Power Measurement Uncertainty per International Guide (Application Note 1449-3), 5988-9215EN for more information on measurement uncertainty.

Typical power accuracy at 25 °C for U2000/1/2H sensors^{13, 14}





 $^{^{13}}$ Measurement uncertainty \leq 1.9%. At room temperature and excluding power level dependent terms (zero set, drift, and noise). Refer to Keysight Fundamentals of RF and Microwave Power Measurements (Part 3) Power Measurement Uncertainty per International Guide (Application Note 1449-3), 5988-9215EN for more information on measurement uncertainty. ¹⁴ After zeroing, 30 minutes of power-on warm-up, and 1024 averages

Zero set, zero drift, and measurement noise

Average only mode

Power range ¹⁵	Zero set (internal)	Zero set (external)	Zero drift ¹⁶	Measurement noise ¹⁷
U2000/1/2A				
–60 to –35 dBm	± 1.5 nW	± 600 pW	200 pW	1 nW
–38 to –15 dBm	± 2 nW	± 1.5 nW	400 pW	1.5 nW
–20 to –6.5 dBm	± 12 nW	± 10 nW	1.5 nW	15 nW
–7.5 to –2 dBm	± 2 μW	± 500 nW	50 nW	650 nW
-4 to 15 dBm	± 4 μW	± 1 µW	500 nW	1 µW
10 to 20 dBm	± 6 µW	± 5 μW	2 µW	10 μW
U2004A				
–60 to –35 dBm	± 2.8 nW	± 600 pW	200 pW	1 nW
–38 to –15 dBm	± 3 nW	± 1.5 nW	400 pW	1.5 nW
–20 to –6.5 dBm	± 12 nW	± 10 nW	1.5 nW	15 nW
–7.5 to –2 dBm	±2μW	± 500 nW	50 nW	650 nW
-4 to 15 dBm	± 4 μW	± 1 µW	500 nW	1 µW
10 to 20 dBm	± 6 µW	± 5 μW	2 µW	10 μW
U2000/1/2H				
–50 to –25 dBm	± 15 nW	± 8 nW	2 nW	10 nW
–28 to –5 dBm	± 20 nW	± 20 nW	4 nW	15 nW
–10 to 3.5 dBm	± 120 nW	± 100 nW	15 nW	150 nW
2.5 to 8 dBm	± 20 μW	± 20 μW	500 nW	6.5 µW
6 to 25 dBm	± 40 μW	± 30 μW	5 µW	10 µW
20 to 30 dBm	± 60 μW	± 60 μW	20 µW	100 μW
U2000/1B sensors				
–30 to –5 dBm	± 1.8 μW	± 800 nW	200 nW	1 µW
–8 to 15 dBm	± 2 μW	± 2 μW	400 nW	1.5 µW
10 to 23.5 dBm	± 12 μW	± 10 μW	1.5 μW	15 μW
22.5 to 28 dBm	± 2 mW	± 1 mW	50 nW	650 μW
26 to 44 dBm	± 4 mW	± 2 mW	500 µW	1 mW

 ¹⁶ Within one hour after zero set, at a constant temperature, after a 24-hour warm-up of the power sensor.
¹⁷ The number of averages at 1 for Normal speed, gate length of 2.27 ms, measured over one-minute interval and two standard deviations.



 $^{^{15}}$ Condition: (i) 0 to 55 °C and (ii) 95% relative humidity at 40 °C non-condensing.

Range ¹⁸	Zero set (internal)	Zero set (external)	Zero drift 19	Measurement noise ²⁰	Noise per sample ²¹
U2000/1/2A					
–38 to –15 dBm	47 nW	43 nW	25 nW	28 nW	90 nW
–20 to –6.5 dBm	530 nW	480 nW	230 nW	300 nW	1 µW
–7.5 to –2 dBm	30 µW	27 µW	19 µW	20 µW	55 µW
–4 to 15 dBm	32 µW	30 µW	24 µW	21 µW	85 μW
10 to 20 dBm	270 µW	200 µW	110 µW	180 μW	550 μW
U2000/1/2H					
–28 to –5 dBm	730 nW	500 nW	300 nW	310 nW	900 nW
–10 to 3.5 dBm	5.3 µW	4.8 µW	3 µW	5 µW	10 μW
–2.5 to 8 dBm	330 µW	270 µW	190 µW	230 µW	550 μW
8 to 25 dBm	440 μW	300 µW	300 µW	260 μW	850 μW
20 to 30 dBm	3.9 µW	2.8 mW	1.1 mW	2.8 mW	5.5 mW
U2000/1B					
–8 to 15 dBm	47 nW	43 µW	25 µW	28 µW	90 µW
10 to 23.5 dBm	530 nW	480 µW	230 µW	300 μW	1 mW
22.5 to 28 dBm	30 µW	27 mW	19 mW	20 mW	55 mW
26 to 44 dBm	32 µW	34 mW	24 mW	21 mW	85 mW

Effect of time-gating and averaging on normal mode measurement noise:

The normal mode measurement noise will depend on the gate length (time-gated period in second) and the number of averages. The noise can be approximately calculated with the following equations.

If the gate length is < 2.73 μ s, use Equation 1:

Noise = Noise per sample
$$\times \frac{1}{\sqrt{Number of averages}}$$

Otherwise, use Equation 2:

Noise = Noise per sample $\times \frac{1}{\sqrt{Number of averages}} \times \left(\frac{1}{Gate \ length/(0.68 \ \mu s)}\right)^{1/4}$

Note: If the noise value obtained from Equation 1 or 2 is lower than the measurement noise specification, use the value as specified in the measurement noise table.

²¹ The Noise Per Sample specification is only applicable for gated power working range stated in the "Normal Mode Key Specifications and Characteristics" table.



¹⁸ Condition: (i) 0 to 55 °C and (ii) 95% relative humidity at 40 °C non-condensing.

¹⁹ Within one hour after zero set, at a constant temperature, after a 24-hour warm-up of the power sensor.

²⁰ The number of averages at 1 for Normal speed, gate length of 2.27 ms, measured over one-minute interval and two standard deviations.

Effects of averaging on noise: Averaging over 1 to 1024 readings is available for reducing noise. The table below provides the measurement noise for a particular sensor with the number of averages set at 16 (for normal mode) and 32 (for x2 mode). Use the noise multiplier, for the appropriate of averages, to determine the total measurement noise value.

Example:

U2000A power sensor, -60 to -35 dBm, normal mode, number of averages = 4

Measurement noise calculation: 1 nW x 1.7 = 1.7 nW

Average only mode

Number of averages	1	2	4	8	16	32	64	128	256	512	1024
Noise multiplier											
Normal mode	2.0	1.8	1.7	1.5	1.0	0.95	0.74	0.55	0.39	0.29	0.21
x2 mode	2.7	2.4	2.0	1.6	1.0	0.91	0.78	0.53	0.34	0.29	0.20

Settling time

Manual filter, 10-dB decreasing power step (not across switching points).

Number of averages	1	2	4	8	16	32	64	128	256	512	1024
Noise multiplier											
Normal mode	2.0	1.8	1.7	1.5	1.0	0.95	0.74	0.55	0.39	0.29	0.21
x2 mode	2.7	2.4	2.0	1.6	1.0	0.91	0.78	0.53	0.34	0.29	0.20



Auto filter, default resolution, 10-dB decreasing power step

	x2 speed	Normal speed	U2000/1/2/4A	U2000/1/2H	U2000/1B	Maximum dBm
1	45 ms	45 ms	— +10 dBm -	+20 dBm	+ + - - +40 dBm	High power path
	82 ms	90 ms				
	1.3 s	2.6 s	— +2 dBm -	+12 dBm	+32 dBm	
Typical settling times	1.5 s	2.7 s	— —4 dBm -	+6 dBm	+26 dBm	Sensor dynamic range
	450 ms	460 ms	— –10 dBm -		+20 dBm	
	1.6 s	2.8 s	— –20 dBm -		+10 dBm	
	20 s	- - 39 s	— –30 dBm -	+ −20 dBm	→ 0 dBm	
	24 s	I		+ −30 dBm ' '	→ −10 dBm ' '	
	24 s	42 s	— –50 dBm -	+ −40 dBm ' '	-+- −20 dBm 	Low power path
					•	Minimum dBm

Settling time with auto filter, default resolution, and a 10 dB decreasing power step (not across the switching point) Settling time = 25 ms^{22}

 $^{^{\}rm 22}$ When a power step crosses through the sensor's auto-range switching point, add 25 ms.



Calibration factor and reflection coefficient

Calibration factor (CF) and reflection coefficient (Rho) data is unique to each sensor. The CF corrects for the frequency response of the sensor. The reflection coefficient (Rho or ρ) relates to the SWR based on the following formula:

$$SWR = \frac{1+\rho}{1+\rho}$$

Maximum relative uncertainties of the CF data are listed in the following table. There is only one set of CF relative uncertainty specification used for both high and low power paths of each sensor.

The uncertainty analysis for the calibration data was done in accordance with the ISO Guide. The uncertainty data reported on the calibration certificate is the expanded uncertainty with a 95% confidence level and a coverage factor of 2.

Frequency	Relative Uncertainty (%) (25 °C ± 10 °C) ²³
U2000A sensor	
10 to 30 MHz	1.8
> 30 MHz to < 2 GHz	1.6
2 to < 14 GHz	2.0
14 to < 16 GHz	2.2
16 to 18 GHz	2.2
U2001A sensor	
10 to 30 MHz	1.8
> 30 MHz to < 2 GHz	1.6
2 to 6 GHz	2.0
U2002A sensor	
50 MHz to 2 GHz	2.0
> 2 to 14 GHz	2.5
> 14 to 16 GHz	2.7
> 16 to 18 GHz	2.7
> 18 to 24 GHz	3.0
U2004A sensor	
9 kHz to 2 GHz	2.0
> 2 to 6 GHz	2.0
U2000B sensor	
10 MHz to 2 GHz	3.0
> 2 to < 12.4 GHz	3.2
12.4 to 18 GHz	3.2

²³ The characterized calibration factor should not deviate between periodic calibrations by more than the specified maximum uncertainty in table. Compliance is confirmed by the relative deviation $\frac{(E_{F_1} - E_{F_1})}{E_{F_1}} = 100$ being less than or equal to $\sqrt{2}$ times the specified maximum uncertainty. $\sqrt{2 * Umax}$ with a reference calibration factor of 100%



Frequency	Relative Uncertainty (%) (25 °C ± 10 °C) ²³	
U2001B sensor		
10 MHz to 2 GHz	3.0	
> 2 to 6 GHz	3.2	
U2000H sensor		
10 MHz to < 8 GHz	2.0	
8 to < 12.4 GHz	2.0	
12.4 to 18 GHz	2.2	
U2001H sensor		
10 MHz to 6 GHz	2.0	
U2002H sensor		
50 MHz to < 8 GHz	2.5	
8 to < 12.4 GHz	2.5	
12.4 to 18 GHz	2.7	
> 18 to 24 GHz	3.0	

Trigger

Internal trigger

Resolution	0.1 dB
Level accuracy	± 1 dB
Jitter	± 1 µs
External TTL trigger input	
Impedance	50 Ω or 1 kΩ ²⁴
Trigger low	< 1.1 V
Trigger high	> 1.9 V
Minimum trigger pulse width	35 ns
Minimum trigger repetition period	80 ns
Trigger latency	11 μs ± 2 μs
Trigger delay	
Range	-0.15 to + 0.15 s
Resolution	1 μs
Trigger hold-off	
Range	1 μs to 400 ms
Resolution	1 μs
Trigger hysteresis	
Range	0 to +3 dB
Resolution	0.1 dB

 $^{^{\}rm 24}\,$ This is only available for option U2001A-H16.

Normal mode key specifications and characteristics

Parameters ²⁵	Performance	
Maximum video bandwidth	40 kHz	
Minimum rise time	40 µs	
Minimum fall time	40 µs	
Range settling time	150 µs	
Minimum pulse width	200 µs	
Sampling rate	1.47 Msps	
Maximum capture length	150 ms	
Maximum pulse repetition rate	150 kHz	
Dynamic range	U2000/1/2A: -30 to +20 dBm	
	U2000/1/2H: -10 to +30 dBm	

U2000/1B: 0 to +44 dBm

General specifications

Dimensions (LxWxH)	U2000/1/4A	163.75 mm x 46.00 mm x 35.90 mm
	U2002A	134.37 mm x 46.00 mm x 35.90 mm
	U2000/1B	308.00 mm x 115.00 mm x 84.00 mm
	U2000/1H	207.00 mm x 46.00 mm x 36.00 mm
	U2002H	164.00 mm x 46.00 mm x 36.00 mm
Neight	U2000/1/4A	0.262 kg
	U2002A	0.226 kg
	U2000/1B	0.762 kg
	U2000/1H	0.324 kg
	U2002H	0.274 kg
Operating environment		
Temperature	0 to 55 °C	
Humidity	Up to 95% relative humidity at 40 °C (non-condensing)	
Altitude	Up to 4600 m (15,000 ft)	
Pollution	Degree 2	
Storage and shipment		
Environment	Sensor should be stored in a clean, dry environment	
Temperature	-30 to +70 °C	
Humidity	Up to 90% relative humidity at 65 °C (non-condensing)	
Altitude	Up to 4600 m (15,000 ft)	
Pollution	Degree 2	

²⁵ Not applicable for U2004A.



Physical characteristics

Other		
Current requirement	200 mA max (approximately)	
Connector	U2000/1/4A, U2000/1H, U2000/1B: N-type (m), 50 Ω	
	U2002A, U2002H: 3.5 mm (m), 50 Ω	
Cable	USB 2.0 Type A to 5-pin Mini-B	
Programmability	SCPI, Keysight VEE, LabVIEW, Microsoft® Visual Basic	
Safety and EMC compliance	IEC 61010-1:2001/EN 61010-1:2001(2nd edition)	
	IEC 61326:2002/EN61326:1997+A1:1998+A2:2001+A3:2003	
	Canada: ICES-001:2004	
	Australia/New Zealand: AS/NZS CISPR11:2004	
Calibration ²⁶	1 year	
Warranty 1	3 years	
Compatible instruments	Keysight handheld spectrum analyzers	
	Keysight signal generators	
	Keysight network analyzers	
	Keysight FieldFox RF/Microwave analyzers	
Mechanical characteristic	Mechanical characteristics such as center conductor protrusion and pin depth are not performance specifications. They are, however, important supplemental characteristics related to electrical performance. At no time should the pin depth of the connector be protruding	

²⁶ See "Ordering information" for available options.



Using the U2000 Series with the BenchVue Software

The U2000 Series is supported by the Keysight BenchVue software's BV0007B Power Meter/Sensor Control and Analysis app. Keysight BenchVue software for the PC accelerates testing by providing intuitive, multiple instrument measurement visibility and data capture with no programming necessary. You can derive answers faster than ever by easily viewing, capturing and exporting measurement data and screen shots. BenchVue software license (BV0007B) is now included with your instrument.



For more information, www.keysight.com/find/BenchVue



Digital meter, analog meter and datalog view.

Supported functionality

Measurement displays	Digital meter		
	Analog meter		
	Data log view		
	Trace view (up to 4 channels or traces on one graph)		
	Multilist with ratio/delta function		
	Compact mode display		
Graph functions	Single marker (up to 5 markers per graph)		
	Dual marker (up to 2 sets of markers per graph)		
	Graph autoscaling		
	Graph zooming		
	Gate measurement analysis (up to 4-pair of gates)		
Pulse characterization functions	17-point automatic pulse parameters characterization		
Instrument settings	Save and recall instrument state including graph settings		
	Instrument preset settings		
	FDO tables		
	Full instrumentation control includes frequency/average/trigger settings, zero, etc.		
Limit and alert function	Sensors Limit and alert notification		
	Alert summary		
Export data or screen shots	Data logging (HDF5/MATLAB/Microsoft Excel/Microsoft Word/CSV)		
	Save screen capture (PNG/JPEG/BMP)		

System and Installation Requirements

PC operating system		
Windows 10, 8 and 7	Windows 10 32-bit and 64-bit (Professional, Enterprise, Education, Home versions)	
	Windows 8 32-bit and 64-bit (Core, Professional, Enterprise)	
	Windows 7 SP1 and later 32-bit and 64-bit (Professional, Enterprise, Ultimate)	
Computer hardware	Professor: 1 GHz or faster (2 GHz or greater recommended)	
	RAM: 1 GB (32-bit) or 2 GB (64-bit) (3 GB or greater recommended)	
Windows XP SP3 32-bit	Processor: 600 MHz or faster (1 GHz or greater recommended)	
(Professional)	RAM: 1 GB (2 GB or greater recommended)	
nterfaces USB, GPIB, LAN, RS-232		
Display resolution	1024 x 768 minimum for single instrument view (higher resolutions are recommended for multiple instrument view)	



Additional requirements

Software: BenchVue requires a VISA (Keysight or National Instruments) when used to connect to physical instruments. Keysight IO Libraries, which contains the necessary VISA, will be installed automatically when BenchVue is installed.

IO Libraries information is available at: www.keysight.com/find/iosuite

Ordering Information

Power sensors

Models	Description	Power range	Connector type
U2000A	10 MHz to 18 GHz USB sensor	–60 to +20 dBm	N-type male, 50 Ω
U2000B	10 MHz to 18 GHz USB sensor	-30 to +44 dBm	N-type male, 50 Ω
U2000H	10 MHz to 18 GHz USB sensor	-50 to +30 dBm	N-type male, 50 Ω
U2001A	10 MHz to 6 GHz USB sensor	-60 to +20 dBm	N-type male, 50 Ω
U2001B	10 MHz to 6 GHz USB sensor	-30 to +44 dBm	N-type male, 50 Ω
U2001H	10 MHz to 6 GHz USB sensor	-50 to +30 dBm	N-type male, 50 Ω
U2002A	50 MHz to 24 GHz USB sensor	-60 to +20 dBm	3.5 mm male, 50 Ω
U2002H	50 MHz to 24 GHz USB sensor	-50 to +30 dBm	3.5 mm male, 50 Ω
U2004A	9 kHz to 6 GHz USB sensor	-60 to +20 dBm	N-type male, 50 Ω
Options and acc	essories		
U2001A-H03	U2001A sensor with extended freque	ncy range, 3 MHz to 6 GH	łz
U2001A-H16	With 1K ohms input trigger impedance. Higher impedance is typically required when several instruments' input trigger ports are connected in parallel for triggering purpose. Standard option has 50 ohms input trigger impedance.		
U2001A-H25	U2001A sensor with extended power range, -60 dBm to +25 dBm		
U2002A-H26	U2002A sensor with extended frequency range, 10 MHz to 26.5 GHz		
Standard-shippe	ed accessories		
Trigger cable BI	NC Male to SMB female 50 Ω , 1.5 m		

Power sensor cable: 1.5 m, 3.0 m, or 5.0 m



Accessories, calibration and documentation options.

Cables	Description		
U2031A	USB 2.0 Type A to 5-pin Mini-B cable with secure locking mechanism, 1.5 m (5 ft)		
U2031B	USB 2.0 Type A to 5-pin Mini-B cable with secure locking mechanism, 3.0 m (10 ft)		
U2031C	USB 2.0 Type A to 5-pin Mini-B cable with secure locking mechanism, 5.0 m (16.4 ft)		
U2032A	BNC (m) to SMB (f) trigger cable, 1.5 m, 50 W		
Travel kits	Description		
U2000A-201	Transit case ²⁷ (500mm (L) x 399mm (W) x 138mm (H))		
U2000A-202	Soft carrying case (405mm (L) x 270mm (W) x 175mm (H))		
U2000B-201	Transit case ²⁸ (500mm (L) x 400mm (W) x 180mm (H))		
U2000A-204	Soft carrying pouch (210mm (L) x 150mm (W) x 70mm (H))		
Hanging kit	Description		
U2000A-203	Holster (104mm (L) x 50mm (W) x 39mm (H))		
Calibration	Description		
Option 1A7	ISO 17025 calibration with test data		
Option A6J	ANSI Z540 calibration with test data		
Software	Description		
BV0007B	BenchVue Power Meter/Sensor Control and Analysis app license		



U2000A/B-201 Transit case



U2000A-202 Soft carrying case



U2000A-203 Holster



U2000A-204 Soft carrying pouch

²⁷ U2000A-201 transit case for A model only.
²⁸ U2000B-201 transit case for B and H model.

Keysight enables innovators to push the boundaries of engineering by quickly solving design, emulation, and test challenges to create the best product experiences. Start your innovation journey at www.keysight.com.



This information is subject to change without notice. © Keysight Technologies, 2011 - 2023, Published in USA, November 27, 2023, 5989-6278EN